

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

## Product Datasheet

### Human Soluble Cluster of Differentiation 40 Ligand (sCD40L) Microsample Fast ELISA Kit BYT-ORB1806755

Artikelname	Human Soluble Cluster of Differentiation 40 Ligand (sCD40L) Microsample Fast ELISA Kit
Artikelnummer	BYT-ORB1806755
Hersteller Artikelnummer	orb1806755
Alternativnummer	BYT-ORB1806755-48,BYT-ORB1806755-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Spezies Reaktivität	Human
Produktbeschreibung	Human Soluble Cluster of Differentiation 40 Ligand (sCD40L) Microsample Fast ELISA Kit...
Detektionsbereich	62.5-4000pg/mL
Sensitivitaet	37.5 pg/mL
UniProt	<a href="#">P29965</a>
Formulierung	Ready to use
Proben	Serum, Plasma and Tissue homogenate and other biological samples,Sample volume:25µL
Target-Kategorie	sCD40L

Anwendungsbeschreibung

Application Notes: This ELISA kit uses the Sandwich-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Human sCD40L. Samples (or Standards) are added to the micro ELISA plate wells and combined with the specific antibody and biotinylated detection antibody specific for Human sCD40L. Then Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Human sCD40L, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm. The OD value is proportional to the concentration of Human sCD40L. You can calculate the concentration of Human sCD40L in the samples by comparing the OD of the samples to the standard curve