

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

## Product Datasheet

### Human Vitamin D Receptor (VDR) ELISA Kit BYT-ORB1807146

Artikelname	Human Vitamin D Receptor (VDR) ELISA Kit
Artikelnummer	BYT-ORB1807146
Hersteller Artikelnummer	orb1807146
Alternativnummer	BYT-ORB1807146-48,BYT-ORB1807146-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Spezies Reaktivität	Human
Produktbeschreibung	Human Vitamin D Receptor (VDR) ELISA Kit...
Detektionsbereich	0.63-40ng/mL
Sensitivitaet	0.251 ng/mL
UniProt	<a href="#">P11473</a>
Proben	serum, Tissue homogenate and Cell lysate
Target-Kategorie	VDR

Anwendungsbeschreibung

Application Notes: This ELISA kit uses the Sandwich-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Human VDR. Standards or samples are added to the micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for Human VDR and Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Human VDR, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm. The OD value is proportional to the concentration of Human VDR. You can calculate the concentration of Human VDR in the samples by comparing the OD of the samples to the standard curve