

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

## Product Datasheet

### Human Tight junction protein ZO-1 (TJP1) ELISA Kit BYT-ORB1807174

Artikelname	Human Tight junction protein ZO-1 (TJP1) ELISA Kit
Artikelnummer	BYT-ORB1807174
Hersteller Artikelnummer	orb1807174
Alternativnummer	BYT-ORB1807174-48,BYT-ORB1807174-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Spezies Reaktivität	Human
Produktbeschreibung	Human Tight junction protein ZO-1 (TJP1) ELISA Kit...
Detektionsbereich	0.16-10ng/mL
Sensitivitaet	0.10 ng/mL
Proben	Tissue homogenate
Target-Kategorie	TJP1

Anwendungsbeschreibung

Application Notes: This ELISA kit uses the Sandwich-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Human TJP1. Standards or samples are added to the micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for Human TJP1 and Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Human TJP1, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm. The OD value is proportional to the concentration of Human TJP1. You can calculate the concentration of Human TJP1 in the samples by comparing the OD of the samples to the standard curve