

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Mouse Beta Crosslaps (beta-CTx) ELISA Kit BYT-ORB1807725

Artikelname	Mouse Beta Crosslaps (beta-CTx) ELISA Kit
Artikelnummer	BYT-ORB1807725
Hersteller Artikelnummer	orb1807725
Alternativnummer	BYT-ORB1807725-48,BYT-ORB1807725-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Spezies Reaktivität	Mouse
Produktbeschreibung	Mouse Beta Crosslaps (beta-CTx) ELISA Kit...
Detektionsbereich	78.13-5000pg/mL
Sensitivitaet	46.88 pg/mL
Proben	serum, plasma, Tissue homogenate and Other biological samples
Target-Kategorie	beta-CTx

Anwendungsbeschreibung

Application Notes: This ELISA kit uses the Competitive-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with Mouse beta-CTx. During the reaction, Mouse beta-CTx in the sample or standard competes with a fixed amount of Mouse beta-CTx on the solid phase supporter for sites on the Biotinylated Detection Ab specific to Mouse beta-CTx. Excess conjugate and unbound sample or standard are washed away, and Avidin-Horseradish Peroxidase (HRP) conjugate are added to each micro plate well and incubated. Then a TMB substrate solution is added to each well. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns from blue to yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm \pm 2 nm. The concentration of Mouse beta-CTx in tested samples can be calculated by comparing the OD of the samples to the standard curve