

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human Cluster of Differentiation 30 Ligand (CD30L) ELISA Kit BYT-ORB1947390

Artikelname	Human Cluster of Differentiation 30 Ligand (CD30L) ELISA Kit
Artikelnummer	BYT-ORB1947390
Hersteller Artikelnummer	orb1947390
Alternativnummer	BYT-ORB1947390-48,BYT-ORB1947390-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Spezies Reaktivität	Human
Produktbeschreibung	Human Cluster of Differentiation 30 Ligand (CD30L) ELISA Kit...
Detektionsbereich	7.82-500pg/mL
Sensitivitaet	4.69 pg/mL
UniProt	P32971
Proben	serum, plasma, Tissue homogenate and Other biological samples
Target-Kategorie	CD30L

Anwendungsbeschreibung

Application Notes: This ELISA kit uses the Sandwich-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Human CD30L. Standards or samples are added to the micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for Human CD30L and Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Human CD30L, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm. The OD value is proportional to the concentration of Human CD30L. You can calculate the concentration of Human CD30L in the samples by comparing the OD of the samples to the standard curve