BIOZOL
Diagnostica Vertrieb GmbH, Oehleckerring 11-13
22419 Hamburg, Germany
Telephone: +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99
E-Mail: info@biozol.de

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Rat beta-Melanocyte Stimulating Hormone (betaMSH) ELISA Kit BYT-ORB1947419

Artikelname	Rat beta-Melanocyte Stimulating Hormone (betaMSH) ELISA Kit
Artikelnummer	BYT-ORB1947419
Hersteller Artikelnummer	orb1947419
Alternativnummer	BYT-ORB1947419-48,BYT-ORB1947419-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Spezies Reaktivität	Rat
Produktbeschreibung	Rat beta-Melanocyte Stimulating Hormone (betaMSH) ELISA Kit...
Detektionsbereich	78.13-5000pg/mL
Sensitivitaet	46.88 pg/mL
Proben	serum, plasma, Tissue homogenate and Other biological samples
Target-Kategorie	betaMSH

Anwendungsbeschreibung	Application Notes: This ELISA kit uses the Sandwich-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Rat betaMSH. Standards or samples are added to the micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for Rat betaMSH and Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Rat betaMSH, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm 2 nm. The OD value is proportional to the concentration of Rat betaMSH. You can calculate the concentration of Rat betaMSH in the samples by comparing the OD of the samples to the standard curve