

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Mouse Betacellulin (BTC) ELISA Kit BYT-ORB1947618

Artikelname	Mouse Betacellulin (BTC) ELISA Kit
Artikelnummer	BYT-ORB1947618
Hersteller Artikelnummer	orb1947618
Alternativnummer	BYT-ORB1947618-48,BYT-ORB1947618-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Spezies Reaktivität	Mouse
Produktbeschreibung	Mouse Betacellulin (BTC) ELISA Kit...
Detektionsbereich	7.82-500pg/mL
Sensitivitaet	4.69 pg/mL
UniProt	Q543J8
Proben	serum, plasma, Tissue homogenate and Other biological samples
Target-Kategorie	BTC

Anwendungsbeschreibung

Application Notes: This ELISA kit uses the Sandwich-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Mouse BTC. Standards or samples are added to the micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for Mouse BTC and Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Mouse BTC, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm. The OD value is proportional to the concentration of Mouse BTC. You can calculate the concentration of Mouse BTC in the samples by comparing the OD of the samples to the standard curve