

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human Thyroid Hormone Receptor Alpha (THRalpha) ELISA Kit BYT-ORB1948091

Artikelname	Human Thyroid Hormone Receptor Alpha (THRalpha) ELISA Kit
Artikelnummer	BYT-ORB1948091
Hersteller Artikelnummer	orb1948091
Alternativnummer	BYT-ORB1948091-48,BYT-ORB1948091-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Spezies Reaktivität	Human
Produktbeschreibung	Human Thyroid Hormone Receptor Alpha (THRalpha) ELISA Kit...
Detektionsbereich	62.5-4000pg/mL
Sensitivitaet	37.5 pg/mL
UniProt	P10827
Proben	Tissue homogenate
Target-Kategorie	THRalpha

Anwendungsbeschreibung

Application Notes: This ELISA kit uses the Sandwich-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Human THRalpha. Standards or samples are added to the micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for Human THRalpha and Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Human THRalpha, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm \pm 2 nm. The OD value is proportional to the concentration of Human THRalpha. You can calculate the concentration of Human THRalpha in the samples by comparing the OD of the samples to the standard curve