BIOZOL
Diagnostica Vertrieb GmbH, Oehleckerring 11-13
22419 Hamburg, Germany
Telephone: +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99
E-Mail: info@biozol.de

Bitte beachten Sie: Dieses Dokument wurde automatisch erstellt und ist kein Ersatz für das Originaldokument des Herstellers.

Product Datasheet

Human Cathepsin C (CTSC) ELISA Kit
BYT-ORB1949952

Artikelname	Human Cathepsin C (CTSC) ELISA Kit
Artikelnummer	BYT-ORB1949952
Hersteller Artikelnummer	orb1949952
Alternativnummer	BYT-ORB1949952-48,BYT-ORB1949952-96
Hersteller	Biorbyt
Kategorie	Kits/Assays
Spezies Reaktivität	Human
Produktbeschreibung	Human Cathepsin C (CTSC) ELISA Kit...
Detektionsbereich	0.32-20ng/mL
Sensitivitaet	0.19 ng/mL
UniProt	P53634
Proben	serum, plasma, Tissue homogenate and Other biological samples
Target-Kategorie	CTSC

Anwendungsbeschreibung	Application Notes: This ELISA kit uses the Sandwich-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Human CTSC. Standards or samples are added to the micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for Human CTSC and Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Human CTSC, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm 2 nm. The OD value is proportional to the concentration of Human CTSC. You can calculate the concentration of Human CTSC in the samples by comparing the OD of the samples to the standard curve