

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Anti-Human IL-15 Antibody, Rabbit, Polyclonal ABT-ABG10210-U050

Article Name	Anti-Human IL-15 Antibody, Rabbit, Polyclonal
Biozol Catalog Number	ABT-ABG10210-U050
Supplier Catalog Number	ABG10210-U050
Alternative Catalog Number	ABT-ABG10210-U050-50UG
Manufacturer	Abcepta
Host	Rabbit
Category	Antikörper
Application	ELISA, IHC, WB
Species Reactivity	Human
Clonality	Polyclonal
Purity	Produced from sera of rabbits pre-immunized with highly pure (>98%) recombinant hIL-15. Anti-Human IL-15 specific antibody was purified by affinity chromatography employing immobilized hIL-15 matrix.
Form	A sterile filtered antibody solution was lyophilized from PBS, pH 7.2.
Antibody Type	Polyclonal Antibody

Application Notes	WesternBlot: To detect hIL-15 by Western Blot analysis this antibody can be used at a concentration of 0.1 - 0.2 µg/ml. Used in conjunction with compatible secondary reagents the detection limit for recombinant hIL-15 is 1.5 - 3.0 ng/lane, under either reducing or non-reducing conditions.. Sandwich: To detect hIL-15 by sandwich ELISA (using 100 µl/well antibody solution) a concentration of 0.5 - 2.0 µg/ml of this antibody is required. This antigen affinity purified antibody, in conjunction with BioGems Biotinylated Anti-Human IL-15 (60-015BT) as a detection antibody, allows the detection of at least 0.2 - 0.4 ng/well of recombinant hIL-15.. Immunohistochemistry: This antibody stained formalin-fixed, paraffin-embedded sections of human normal placenta. The recommended concentration is 0.5 mg/ml-1.0 mg/ml with a two hour incubation at room temperature. An HRP-labeled polymer detection system was used with a DAB chromogen. Heat induced antigen retrieval with a pH 6.0 sodium citrate buffer is recommended. Optimal concentrations and conditions may vary.. Neutralization: To yield one-half maximal inhibition [ND50] of the biological activity of hIL-15 (1.0 ng/ml), a concentration of 0.15 - 0.20 µg/ml of this antibody is required.. Reconstitution: Centrifuge vial prior to opening. Reconstitute in sterile water to a concentration of 0.1-1.0 mg/ml.
-------------------	--