

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

Product Datasheet

Human AMA-M2(Anti-Mitochondrial Antibody M2 Subtype) ELISA Kit EBT-ELK0813

Article Name	Human AMA-M2(Anti-Mitochondrial Antibody M2 Subtype) ELISA Kit
Biozol Catalog Number	EBT-ELK0813
Supplier Catalog Number	ELK0813
Alternative Catalog Number	EBT-ELK0813-96, EBT-ELK0813-48, EBT-ELK0813-96X5
Manufacturer	ELK Biotechnology
Category	Kits/Assays
Species Reactivity	Human
Concentration	20 ng/mL
Range	0.32-20 ng/mL
Sensitivity	0.13 ng/mL
Samples	serum, plasma and other biological fluids

Application Notes	Assay Type: Sandwich. Assay length: 3.5h. Research Area: Infection immunity,. Test principle: The test principle applied in this kit is Sandwich enzyme immunoassay. The microtiter plate provided in this kit has been pre-coated with an antibody specific to Human AMA-M2. Standards or samples are added to the appropriate microtiter plate wells then with a biotin-conjugated antibody specific to Human AMA-M2. Next, Avidin conjugated to Horseradish Peroxidase (HRP) is added to each microplate well and incubated. After TMB substrate solution is added, only those wells that contain Human AMA-M2, biotin-conjugated antibody and enzyme-conjugated Avidin will exhibit a change in color. The enzyme-substrate reaction is terminated by the addition of sulphuric acid solution and the color change is measured spectrophotometrically at a wavelength of 450nm 10nm. The concentration of Human AMA-M2 in the samples is then determined by comparing the OD of the samples to the standard curve
-------------------	--