

Please note: This document was created automatically and is not a substitute for the manufacturer's original document.

## Product Datasheet

### Human A Disintegrin and Metalloproteinase with Thrombospondin 12 (ADAMTS12) ELISA Kit BYT-ORB1947863

Article Name	Human A Disintegrin and Metalloproteinase with Thrombospondin 12 (ADAMTS12) ELISA Kit
Biozol Catalog Number	BYT-ORB1947863
Supplier Catalog Number	orb1947863
Alternative Catalog Number	BYT-ORB1947863-48,BYT-ORB1947863-96
Manufacturer	Biorbyt
Category	Kits/Assays
Species Reactivity	Human
Product Description	Human A Disintegrin and Metalloproteinase with Thrombospondin 12 (ADAMTS12) ELISA Kit...
Range	78.13-5000pg/mL
Sensitivity	46.88 pg/mL
UniProt	<a href="#">P58397</a>
Samples	serum, plasma, Tissue homogenate and Other biological samples
Target	ADAMTS12

Application Notes

Application Notes: This ELISA kit uses the Sandwich-ELISA principle. The micro ELISA plate provided in this kit has been pre-coated with an antibody specific to Human ADAMTS12. Standards or samples are added to the micro ELISA plate wells and combined with the specific antibody. Then a biotinylated detection antibody specific for Human ADAMTS12 and Avidin-Horseradish Peroxidase (HRP) conjugate are added successively to each micro plate well and incubated. Free components are washed away. The substrate solution is added to each well. Only those wells that contain Human ADAMTS12, biotinylated detection antibody and Avidin-HRP conjugate will appear blue in color. The enzyme-substrate reaction is terminated by the addition of stop solution and the color turns yellow. The optical density (OD) is measured spectrophotometrically at a wavelength of 450 nm  $\pm$  2 nm. The OD value is proportional to the concentration of Human ADAMTS12. You can calculate the concentration of Human ADAMTS12 in the samples by comparing the OD of the samples to the standard curve