BIOZOL
Diagnostica Vertrieb GmbH, Oehleckerring 11-13
22419 Hamburg, Germany
Telephone: +49 (0)89 3799666-6 | **Fax:** +49 (0)89 3799666-99
E-Mail: info@biozol.de

Product Datasheet

Swine anti Human Hepatitis B associated antigen, Clone: [Polyclonal], Porcine, Monoclonal
NMB-SWAHU/HAA

Article Name	Swine anti Human Hepatitis B associated antigen, Clone: [Polyclonal], Porcine, Monoclonal
Biozol Catalog Number	NMB-SWAHU/HAA
Supplier Catalog Number	SwAHu/HAA
Alternative Catalog Number	NMB-SWAHU/HAA
Manufacturer	NordicMubio
Host	Porcine
Category	Antikörper
Product Description	The defined antibody specificity is to the hepatitis associated antigen as tested in counterelectrophoresis and double radial immunodiffusion against HAA positive human serum. The product does not react with normal human serum or plasma. In precipita...
Clonality	Monoclonal
Clone Designation	[Polyclonal]
Buffer	Delipidated, heat inactivated, lyophilized, stable whole serum. No preservative added. Total protein and IgG concentrations in the antiserum are comparable to those of normal pooled swine serum. No foreign proteins added. Reconstitute the lyophilized anti
Source	The immunogen has been prepared from sera of patients with hepatitis using gel filtration Freunds complete adjuvant is used in the first step of the immunization procedure.

Formula	Delipidated, heat inactivated, lyophilized, stable whole serum. No preservative added. Total protein and IgG concentrations in the antiserum are comparable to those of normal pooled swine serum. No foreign proteins added.
Application Notes	Precipitation assays. In counter electrophoresis use opposite holes at a distance of 4mm with 6 µl antiserum and serum samples. In double radial immunodiffusion use a rosette arrangement with 10 µl antiserum in 3 mm diameter center well and 2 µl plasma samples (neat and serially diluted) in 2 mm diameter peripheral wells. The distance between the wells should be 2-3 mm. In case the concentration of the protein component is significantly increased or reduced, the required equivalence between the reactants can be establish by increasing the amount of antiserum or antigen respectively.